

The Summer Undergraduate Research Fellowship (SURF) Symposium
7 August 2014
Purdue University, West Lafayette, Indiana, USA

Validating Hemoglobin Saturation and Dissolved Oxygen in Tumors using Photoacoustic Imaging and the OxyLab Probe

Jenna Burnett¹, Justin Sick², Marc Mendonca³, Keith Stantz²

¹Department of Physics and Astronomy, Purdue University

²School of Human and Health Science, Purdue University

³School of Medicine, Indiana University

ABSTRACT

The goal of this experiment is to validate the relationship between hemoglobin saturation (SaO_2) and partial pressure of dissolved oxygen (pO_2) in breast tumors in mice using photoacoustic computed tomographic (PCT) imaging and OxyLite probe, respectively. In its simplest form, the relationship is described by the dissociation curve, or Hill's equation, for hemoglobin, and is modeled as a sigmoidal curve that is a function of two parameters – the Hill coefficient, n , and the net association constant of HbO_2 , K (or pO_2 at 50% SaO_2). First, a calibration study to validate Hill's equation in blood was performed by creating a closed circuit phantom to test the SaO_2 (co-oximeter) and pO_2 (Oxylite probe) relationship ($K=23.2\text{mmHg}$ and $n=2.26$). Next, non-invasive localized measurements of SaO_2 in MDA-MD-231 and MCF7 breast tumors using PCT spectroscopic methods were compared to pO_2 levels, where pO_2 levels were measured in 1mm increments across the central axis of the tumor. The fitted results for MCF7 and MDA-MD-231 were $K=17.2\text{mmHg}$ and 20.7mmHg , and $n=1.76$ and 1.63 , respectively. The results are consistent with sigmoidal form of Hill's equation. The lower value of K is indicative of the acidic microenvironment associated with tumors. Ongoing work to correct for photon transport and image artifacts are anticipated to enhance the quality of the results. In conclusion, the results from this study demonstrate photoacoustic can be used to measure tumor oxygenation, and its potential use in investigating the effectiveness of anti-angiogenesis therapy.

KEYWORDS

Hemoglobin saturation, partial pressure, dissolved oxygen, oxygen probe, photoacoustic, imaging, breast tumors, MCF7, MDA-MD231

REFERENCES

- Zhang, Y., Hong, H., & Cai, W., 2010. Photoacoustic Imaging. Cold Spring Harbor Protocol.
- Leow, M.K.S. 2007. Configuration of the hemoglobin oxygen dissociation curve demystified: a basic mathematical proof for medical and biological sciences undergraduates. *Advan in Physiol Edu.* 31: 198-201.
- Pittman, R.N. 2011. Regulation of tissue oxygenation. *Morgan and Claypool Life Sciences, CA.* Ch. 10